

Addendum for CMR-3241

UltraScan VIS Sample Clamp Modified with Spring

Instruments: UltraScan VIS

Description: This CMR replaces the standard sample clamp to include a spring to provide constant tension.

Note: Modification of sample clamp includes removal of rear activation button to allow sample clamp to have full action and constant tension.



Installation:

- 1- Remove the existing sample clamp and install the modified sample clamp.
- 2- Proceed with standardization and sample readings.