



**Certificate for Haze Standards with Values Read at Factory
with HunterLab and Haze Holder**

Date:

Room Temperature: 23° +/- 2° C

Catalogue Number	Haze Standard Serial Number	Haze Standard	ASTM D1003 Assigned Haze Values	HunterLab with CMR2760 Haze Holder Haze Values Read at Factory C/2
4740		H1	0.85 +/- 0.26	
4741		H5	3.72 +/- 0.26	
4742		H10	9.26 +/- 0.26	
4743		H20	19.40 +/- 0.26	
4744		H30	34.00 +/- 0.26	

Bill Babiarez
Quality Assurance Manager



Hunter Associates Laboratory, Inc.

11491 Sunset Hills Road
Reston, Virginia USA 20190
NET: www.hunterlab.com
TEL: 703-471-2176