


## Addendum to User's Manual

---

### Description

This CMR modification incorporates a spill tray with a standard C02-1005-481 transmission cell holder that is used with all standard HunterLab transmission cells (extra).



### Installation

Align the holder against the TTRAN port on the instrument and tighten the screws to secure the device to the transmission floor.

### Operation

This device is designed to fit tightly at the TTRAN port of all HunterLab sphere instruments minimizing potential damage should a spill occur.